



SPECIFICATION SHEET

UltraVision Large Volume Detection System Anti-Rabbit, HRP (Ready-To-Use)

DESCRIPTION:

The reagents in this kit constitute a labeled streptavidin-biotin immunoenzymatic antigen detection system. This technique involves the sequential incubation of the specimen with an unconjugated primary antibody specific to the target antigen, a biotinylated secondary antibody which reacts with the primary antibody, enzyme-labeled streptavidin, and substrate-chromogen.

SPECIES OF ORIGIN:	Goat						
ANTIGEN SPECIFICITY:	Anti-Rabbit IgG (H+L)						
PREADSORBED AGAINST:	Human, Mouse, Rat						
ENZYME CONJUGATE:	Peroxidase						
CHROMOGEN/SUBSTRATE:	None provided						
AVAILABILITY:	<table><thead><tr><th><u>Catalog #</u></th><th><u>Slide Volume</u></th></tr></thead><tbody><tr><td>TR-060-HL</td><td>600-1200 slides</td></tr><tr><td>TR-125-HL</td><td>1250-2500 slides</td></tr></tbody></table>	<u>Catalog #</u>	<u>Slide Volume</u>	TR-060-HL	600-1200 slides	TR-125-HL	1250-2500 slides
<u>Catalog #</u>	<u>Slide Volume</u>						
TR-060-HL	600-1200 slides						
TR-125-HL	1250-2500 slides						

STAINING PROTOCOL (kit components in bold):

1. Deparaffinize and rehydrate tissue section.
2. To reduce nonspecific background staining due to endogenous peroxidase, incubate slide in hydrogen peroxide for 10-15 minutes.
3. Wash 2 times in buffer.
4. If required, incubate tissue in digestive enzyme (or appropriate pretreatment).
5. Wash 4 times in buffer.
6. (Optional) Apply **Large Volume Ultra V Block** and incubate for 5 minutes at room temperature to block nonspecific background staining. **Note:** Do not exceed 10 minutes or there may be a reduction in desired stain.

47790 Westinghouse Dr. Fremont CA 94539 USA
Tel: (800) 828-1628 Fax: (510) 991-2826
Email: labvision@labvision.com Website: www.labvision.com



TR-125-HL
Rev 21211
Page 1 of 2



7. Rinse (Optional).
8. Apply primary antibody and incubate according to manufacturer's protocol.
9. Wash 4 times in buffer.
10. Apply **Large Volume Biotinylated Goat Anti-Rabbit** and incubate for 10 minutes at room temperature.
11. Wash 4 times in buffer.
12. Apply **Large Volume Streptavidin Peroxidase** and incubate for 10 minutes at room temperature.
13. Rinse 4 times in buffer.
14. Incubate with peroxidase-compatible chromogen of choice according to manufacturer's recommendations.
15. Counterstain and coverslip.

STORAGE: Store at 2-8°C.

SPECIFICATION SHEET

TROUBLESHOOTING GUIDE

OVERSTAINING:

1. Concentration of the primary antibody was too high or the incubation time was too long.
2. Temperature during incubation was too high.
3. Incubation time with link antibody or streptavidin/enzyme label was too long.

NONSPECIFIC BACKGROUND STAINING:

1. Rinsing between steps was inadequate.
2. Tissue was allowed to dry with reagents on.
3. Folds in tissue trapped reagents.
4. Tissue contains endogenous peroxidase.
5. Tissue contains endogenous biotin.
6. Antigen migrated in tissue.
7. Excessive tissue adhesive on slides.
8. Inadequate blocking with protein block.

WEAK STAINING:

47790 Westinghouse Dr. Fremont CA 94539 USA
Tel: (800) 828-1628 Fax: (510) 991-2826
Email: labvision@labvision.com Website: www.labvision.com



TR-125-HL
Rev 21211
Page 2 Of 2



1. Primary antibody concentration was too low or incubation time was too short.
2. Reagents are past their expiration date.
3. Inadequate removal of wash buffer between steps, resulting in dilution of reagents.
4. Counterstain or mounting media were incompatible and dissolved the chromogen reaction product.
5. Room temperature was excessively cool.
6. The primary antibody does not recognize an antigen that survives fixation and embedding in high enough amounts.
7. Excessive incubation with protein block (Ultra Block or normal serum).

NO STAINING:

1. Steps were inadvertently left out.
2. There is no antigen in the tissue.
3. The primary antibody is not of rabbit origin.
4. Chromogenic substrate has been replaced with another that is not intended for use with peroxidase.
5. One or more components of the kit have been inactivated.

LIMITATIONS:

This product is available for research use only; not for diagnostic or therapeutic work. LabVision Corporation will not be held responsible for patent infringement or other violation that may occur with the use of this product.