

CD3 (Early T Cell Marker)

Catalog # RM-9107-S0, -S1, or -S (0.1ml, 0.5ml, or 1.0ml)

Catalog # RM-9107-R7 (7.0ml)

Catalog # RM-9107-RQ (12.0ml)

Please note this data sheet has been changed effective June 7, 2011

INTENDED USE:

- **For In Vitro Diagnostic Use:** This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded tissue sections, to be viewed by light microscopy.
- **Description:** This antibody reacts with the intracytoplasmic portion of the CD3 antigen expressed by T cells. It stains human T cells in both the cortex and medulla of the thymus and in peripheral lymphoid tissues. This antibody is suitable for staining normal and neoplastic T cells in formalin-fixed, paraffin-embedded tissues.
- **Expected Staining Pattern:** Cell Membrane
- **Positive Control:** Jurkat cells, Tonsil

MATERIALS PROVIDED:

CD3 (Early T Cell Marker) (refer to catalog number):

- #RM-9107-S (or -S0, -S1): Tissue culture supernatant, concentrated, with 0.09% Sodium Azide.
or
- #RM-9107-R7: (7.0ml) of antibody prediluted in 0.05mol/L Tris-HCl, pH 7.6 containing stabilizing protein and 0.015mol/L sodium azide.
or
- #RM-9107-RQ: (12.0ml) of antibody prediluted in 0.025mol/L Tris/Tris-HCl, pH 7.4 ±0.1 containing stabilizing protein and 0.015mol/L sodium azide.
- **Antibody Concentration:** Not known
- **Host:** Rabbit
- **Epitope:** 156-168aa
- **Species Reactivity:** Human. Other species not tested.
- **Clone Designation:** SP7
- **Ig Isotype / Light Chain:** Rabbit IgG
- **Immunogen:** A synthetic 13-mer peptide corresponding to aa 156-168 of the epsilon chain of human CD3 protein.
- **Microbiological State:** This product is not sterile.

MATERIALS REQUIRED, BUT NOT PROVIDED:

- **Antibody Diluent:** For concentrated antibodies, the antibody must be diluted before using. Use Lab Vision Antibody Diluent (catalog # TA-125-UD). Refer to diluent product instructions for use.
- **Negative Control Reagent:** Refer to the "General Protocol" instructions.
- **Visualization System:** Refer to the "General Protocol" instructions.

METHODS AND PROCEDURES:

Using UltraVision LP detection systems	
Specimen Preparation	Refer to the "General Protocol" instructions.
Dilution of Concentrated Antibody	1:150 in antibody diluent
Tissue Section Pretreatment	Staining of formalin-fixed tissue sections requires treating the tissue sections in boiling 10mM citrate buffer, pH 6.0 (Lab Vision catalog # AP-9003), for 10-20 minutes followed by cooling at room temperature for 20 min.
Primary Antibody Incubation Time	20 minutes at Room Temperature
Visualization	To detect antibody, follow the instructions provided with the visualization system.

CD3 (Early T Cell Marker)

Catalog # RM-9107-S0, -S1, or -S (0.1ml, 0.5ml, or 1.0ml)

Catalog # RM-9107-R7 (7.0ml)

Catalog # RM-9107-RQ (12.0ml)

Please note this data sheet has been changed effective June 7, 2011

Using UltraVision Quanto detection systems	
Specimen Preparation	Refer to the "General Protocol" instructions.
Dilution of Concentrated Antibody	1:300 in antibody diluent
Tissue Section Pretreatment	Staining of formalin-fixed tissue sections requires treating the tissue sections in boiling 10mM citrate buffer, pH 6.0 (Lab Vision catalog # AP-9003), for 10-20 minutes followed by cooling at room temperature for 20 min.
Primary Antibody Incubation Time	20 minutes using the Quanto detection systems at Room Temperature
Visualization	To detect antibody, follow the instructions provided with the visualization system.

STORAGE and STABILITY:

This product contains sodium azide and is stable for 24 months when stored at 2-8°C. Do not use after expiration date indicated on label of the product. If reagent is not stored as recommended, performance must be validated by the user.